



CERTIFICATE OF ANALYSIS

Version 3.7

revised on June 7, 2023

Glycan Sequencing Kit

PRODUCT NAME: Glycan Sequencing Kit
PRODUCT NUMBERS: KE-SQ01

LOT NUMBER: 101.1B
EXPIRATION DATE: June 3, 2026
ASSAY DATE: May 27, 2023

STORAGE: 4°C (Do not freeze)

ENZYME: Neuraminidase Au
PARENT LOT NUMBER: 201.2A
SOURCE: recombinant gene from *Arthrobacter ureafaciens* expressed in *E. Coli*
FORMULATION: 20 mM Tris-HCl, 20 mM NaCl pH 7.5, filter sterilized
ACTIVITY: Activity¹ 5 U/ml (Specification: ≥ 5 U/ml)
Specific Activity² 210 U/mg (Specification: ≥ 135 U/mg)

CONTAMINATING ACTIVITY ASSAYS:
Protease Assay³ Passed (Specification "Passed")

ENZYME: β(1-4) Galactosidase
PARENT LOT NUMBER: 505.1C
SOURCE: recombinant gene from *Streptococcus pneumoniae* expressed in *E. Coli*
FORMULATION: 20 mM Tris-HCl, 25 mM NaCl pH 7.5, filter sterilized
ACTIVITY: Activity¹ 3 U/ml (Specification: ≥ 3 U/ml)
Specific Activity² 12.4 U/mg (Specification: ≥ 6 U/mg)

CONTAMINATING ACTIVITY ASSAYS:
Protease Assay³ Passed (Specification "Passed")

ENZYME: N-Acetylglucosaminidase
PARENT LOT NUMBER: 502.1A
SOURCE: recombinant gene from *Streptococcus pneumoniae* expressed in *E. Coli*
FORMULATION: 20 mM Tris-HCl pH 7.5, 25 mM NaCl filter sterilized
ACTIVITY: Activity¹ 50 U/ml (Specification: ≥ 40 U/ml)
Specific Activity² 118 U/mg (Specification: ≥ 80 U/mg)

CONTAMINATING ACTIVITY ASSAYS:
Protease Assay³ Passed (Specification "Passed")

ENZYME: Alpha-(1-2,3,6) Mannosidase
PARENT LOT NUMBER: 401.1G
SOURCE: Jack Bean
FORMULATION: 20 mM Tris pH 7.5, 50 mM NaCl, 0.1 mM zinc chloride, filter sterilized
ACTIVITY: Activity¹ 20 U/ml (Specification: ≥ 9 U/ml)
Specific Activity² 7.1 U/mg (Specification: ≥ 4 U/mg)

CONTAMINATING ACTIVITY ASSAYS:
Protease Assay³ Passed (Specification "Passed")
Glycosidase activity⁴
Beta-Galactosidase Passed (Specification "Passed")
N-acetylglucosaminidase Passed (Specification "Passed")
Alpha-Galactosidase Passed (Specification "Passed")

(Continued on page 2)



CERTIFICATE OF ANALYSIS

Version 3.7

revised on June 7, 2023

Glycan Sequencing Kit

(Page 2 of 2)

PRODUCT NAME: Glycan Sequencing Kit
PRODUCT NUMBER: KE-SQ01

LOT NUMBER: 807.1A
EXPIRATION DATE: June 3, 2026
ASSAY DATE: May 27, 2023

STORAGE: 4°C (Do not freeze)

ENZYME: $\alpha(1-6)$ Core Mannosidase
PARENT LOT NUMBER: 126.1A
SOURCE: recombinant gene from *Streptococcus pneumoniae* expressed in *E. Coli*
FORMULATION: 50 mM sodium phosphate pH 7.5, 0.1 mM zinc chloride, filter sterilized
ACTIVITY: Activity¹ 1 U/ml (Specification: ≥ 1 U/ml)
Specific Activity² 1.2 U/mg (Specification: ≥ 0.75 U/mg)
CONTAMINATING ACTIVITY ASSAYS: Protease Assay³ Passed (Specification "Passed")

1. Defined as the amount of enzyme required to hydrolyze 1 μ mole of p-nitrophenyl-alpha-p-mannoside to p-nitrophenol in 1 minute at pH 5.0 and 37°C.
2. Protein concentration determined by Bradford method, using BSA as a standard.
3. For the protease assay, 10 μ g of denatured BSA is incubated at 37°C for 24 hr with 2 ul of enzyme. SDS-PAGE analysis of the treated BSA shows no evidence of degradation.
4. Unexpected glycosidic activity is tested by incubating the enzyme for 24 hours at 37°C with the appropriate substrates; the detection limit of this assay is 5 μ U/ml (IUB). A passing lot will have no detectable activity.

President - QA-Bio Inc
Authorized by Mike Gibson
June 7, 2023